

From Hospital To Home

The Vivo 55 is an advanced ventilator designed to deliver secure and comfortable treatment to patients from 10 kg. The Vivo 55 can be used for a wide variety of patients, both life-support and nonlife-support, thanks to a comprehensive set of modes, circuits and accessories.

The extensive monitoring capabilities help to obtain a better insight into the quality of ventilation. The Vivo 55 is an excellent choice for mechanical ventilation in hospital and long-term care facilities. The Vivo 55 is prepared for connection to Breas cloud solutions.

PERFORMANCE

- Highly accurate volume delivery and trigger sensitivity
- Comprehensive set of modes, including mouthpiece ventilation and SIMV
- Wide range of settings and alarms to adapt to individual patient's need
- Invasive and non-invasive ventilation; circuits with active exhalation valve and leakage port
- Intuitive user interface with direct-access buttons
- Low noise level



EXTENSIVE MONITORING

- Optional Intergrated: SpO_2, CO_2 and Fio_2, Monitoring
- Numerical data, wave forms and trends on the display
- PC software is enabling real-time monitoring, wave forms, loops, 365 days of data storage and detailed reports
- Prepared for connection to Breas cloud solutions*

FLEXIBILITY

- Robust Scandinavian design for hospital and mobile use
- Wide range of accessories for hospital use
- 12-hrs autonomy with the 4-hrs internal battery and 8-hrs optional Click-on battery combined
- Optional Protective cover for safe outdoor use
- 3 Profiles to personalize treatment according to the patient's requirements

*Subjected to Availablity

BREAS

Vivo 55 Technical Specifications

Settings / Performance	
Ventilation Modes	PSV, PSV(TgV), PCV SIMV, PCV, PCV(TgV), VCV SIMV PCV(A), PCV(A+TgV), CPAP, VCV, VCV(A), VCV MPV, PCV MPV
Device Modes	Home, Clinical
Inspiratory Pressure	4 to 50 cmH_2O
PEEP	Off, 2 cmH2O to 30 cmH_2O
Breath Rate (PCV, VCV, MPV, SIMV)	4-40 bpm, 0-40 bpm in MPV mode
Inspiratory Time	0.3 to 5 s
Backup Inspiratory Time	0.3 to 5s (PSV)
Rise Time	Auto, 1 to 9 (PSV & PCV) 50-90 %, Off (VCV)
Inspiratory Trigger	Auto, 1 to 9 (PSV, PCV & VCV), Off (PCV & VCV)
Expiratory Trigger	Auto, 1 to 9 (PSV)
Minimum Inspiratory Time	Off, 0.3 to 3 s
Maximum Inspiratory Time	0.3 to 3 s, Off
Target Volume	100-2500 ml
Tidal Volume	100-2500 ml
Flow Pattern	Square, decelerating
Sigh Function	On/Off, rate (every 50-100-150-200-250 breaths), sigh% (125,150,175,200%)
Monitoring	
PC Software	Pressure flow Volume loops and graphs for therapy live Monitoring and Recording.
Displayed data	Ppeak, PEEP, Pmean, Leakage, MVe/MVI, Vte/Vti, FiO_2, % in TgV, Total Rate, Spont Rate, % Spont, SpO_2, Pulse Rate, $EtCO_2$, $InspCO_2$, I:E, InspTime, Rise Time
Waveforms	Pressure, Flow, Volume
Trends over 1, 6, 24 and 48 h	Ppeak, PEEP, Total rate, Spont rate, Vt, Leakage, SpO_2, $EtCO_2$
Power Supplies	
Mains supply	100 to 240 V AC
External DC	24 V DC
Optional Click-on battery	8 hours
Internal battery	4 hours
XPAC External Battery (Optional)	Input 100-240v,Output 18v DC, 60w Max
Main alarms	
Alarms	High Pressure, Low Pressure, High PEEP, Low PEEP, High Vte/Vti, Low Vte/Vti, High MVe/MVi, Low MVe/MVi, High Breath Rate, Low Breath Rate, Apnea, Disconnection, Rebreath- ing, High FiO_2, Low FiO_2, High SpO_2, Low SpO_2, High $EtCO_2$, Low $EtCO_2$, High $InspCO_2$,High Pulse Rate, Low Pulse Rate, Low Last Power Source.
Dimensions	
WxHxD	343 x 125 x 264 mm (343 x 125 x 285 mm with click-on battery)
Weight	5Kg
Noise level (at 10 cmH2O constant pressure)	Less than 30 dB(A)

Intended use:

The Vivo 55 ventilator (with or without the SpO2 and CO2 sensor) is intended to provide continuous or intermittent ventilatory support for the care of individuals who require mechanical ventilation. Specifically, the ventilator is applicable for pediatric through adult patients weighing more than 10kg or 22lbs. The Vivo 55 with the SpO2 sensor is intended to measure functional oxygen saturation of arterial hemoglobin (%SpO2) and pulse rate. The Vivo 55 with the CO2 sensor is intended to measure CO_2 in the inspiratory and expiratory gas. The device is intended to be used in home, institution, hospitals and portable applications such as wheelchairs and gurneys. It may be used for both invasive and non-invasive ventilation.

Vivo by Breas is a trademark of Breas Medical AB. 2019 Breas Medical - All rights reserved. Breas Medical reserves the right to make changes in specifications and features shown herein, or discontinue the product described at any time without notice or obligation. Contact your Breas representative for the most current information. Breas and the Breas logo are trademarks of Breas Medical AB. COR-SAL-MAR-000362-Rev1

Företagsvägen 1-SE-435 33 Mölnlycke Sweden
Phone +46 31 86 88 00
www.breas.com

BREAS